

*EG-Konformitätserklärung*

*EC Declaration of Conformity*

Die Lichtquelle:

The light source:

**KL 1500 LCD**

erfüllt die Bestimmungen folgender Richtlinien des Rates der Europäischen Gemeinschaft:

- 2004/108/EG (EMV-Richtlinie)
- 2006/95/EG (Niederspannungs-Richtlinie)
- 93/42/EWG (Medizingeräte-Richtlinie)

Die Lichtquellen sind für den Betrieb mit einer Nennspannung von 120V bzw. 230V ausgelegt.

Die Übereinstimmung der Lichtquellen mit den wesentlichen Schutzanforderungen obiger Richtlinien wird durch die vollständige Einhaltung folgender Normen nachgewiesen:

is in conformity with the following European Directives:

- 2004/108/EC (EMC Directive)
- 2006/95/EG (Low Voltage Directive)
- 93/42/ECC (Medical Devices Directive)

The light sources are designed for use with a nominal voltage of 120V or 230V.

Full compliance with the standards listed below proves the conformity of the light sources with the essential protection requirements of the above mentioned EC directives:

- |   |  |
|---|--|
| <ul style="list-style-type: none"><li>• EN 55011:2007+A1:2004+A2:2007</li><li>• EN 61000-3-2:2006</li><li>• EN 61000-3-3:1995+A1:2001+A2:2005</li><li>• EN 61326-1:2006</li></ul> | <ul style="list-style-type: none"><li>• EN 61010-1:2001</li><li>• DIN EN 60601-1 (VDE 0750 Teil 1):1996-03<br/>EN 60601-1:1990 + A1:1993 + A2:1995</li><li>• DIN EN 60601-1-2 (VDE 0750 Teil 1-2) :2002-10<br/>EN 60601-1-2:2001</li></ul> |
|---|--|

Die Lichtquelle trägt das CE-Konformitätskennzeichen

The light source bears the CE conformity mark



Mainz, November 2009

ppa.

Dr. Burkhard Danielzik

Business Unit Lighting and Imaging  
Vice President Europe

i.V.

Jörg Warrelmann

Business Unit Lighting and Imaging  
Business Manager Industrial & Medical